

Tech-Cal

HEMATOLOGY CALIBRATOR

CALIBRATOR



LOT

T0222



2022-03-05

Calibration Values for ADVIA[®] 120, 2120 & 2120i Hematology Systems

Parameter	Calibration Values	
WBCB × 10 ³ /μL (BASO)	7.45	
WBCP × 10 ³ /μL (PEROX)	7.80	
RBC × 10 ⁶ /μL	4.58	
HGB g/dL	13.3	
MCV fL	78.0	
CHCM %	38.5	
PLT × 10 ³ /μL	254	
NEUT _x %	59.0	± 6.0
NEUT _y %	75.5	± 5.0



Before using, refer to the instruction sheet for mixing directions.

For technical assistance in the USA and Canada call Technical Service at (800) 523-3395.

All brands and products are trademarks or registered trademarks of their respective companies.