

Tech-Cal

HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT

T0523



2023-06-05

Calibration Values for ADVIA[®] 120, 2120 & 2120i Hematology Systems

Parameter	Calibration Values
WBCB × 10 ³ /μL (BASO)	7.70
WBCP × 10 ³ /μL (PEROX)	8.10
RBC × 10 ⁶ /μL	4.46
HGB g/dL	13.4
MCV fL	78.5
CHCM %	39.5
PLT × 10 ³ /μL	215
NEUT _x %	62.0 ± 6.0
NEUT _y %	77.5 ± 5.0



Before using, refer to the instruction sheet for mixing directions.

For technical assistance in the USA and Canada call Technical Service at (800) 523-3395.

All brands and products are trademarks or registered trademarks of their respective companies.